



ATTESTATION OF CONFORMITY

The following product has been tested by us with the listed standards and found in conformity with the council EMC Directive 2014/30/EU. It is possible to use CE marking to demonstrate the conformity with this EMC Directive.

<i>Manufacture's Name and Address:</i>	<i>ACCU-SCOPE Inc 73 Mall Drive Commack, New York 11725</i>
<i>Product Identification:</i>	<i>EXI-410 Series Microscopes EXI-410 Series Fluorescence Microscopes</i>
<i>Description:</i>	<i>Laboratory Clinical Research Microscope</i>
<i>Trade Name:</i>	<i>ACCU-SCOPE Inc</i>
<i>Model Number:</i>	<i>EXI-410 Series Microscopes</i>
<i>Identification of Test Standards:</i>	<i>EN 61326-1:2013 EN 61000-3-2:2014 EN 61000-3-3:2013</i>
<i>Report No.:</i>	<i>EN180620011E</i>
<i>Testing Facilities:</i>	<i>EMTEK</i>

The holder of this Certificate is authorized to use in connection with CE declaration of conformity with this EMC Directive.

Year of CE Markings: June 2018

A handwritten signature in blue ink, appearing to read 'H Singh'.

*Heston Singh
VP, Business Development
Legally binding signature of the issuer*