



## ***ATTESTATION OF CONFORMITY***

*The following product has been tested by us with the listed standards and found in conformity with the council LVD Directive 2014/35/EU. It is possible to use CE marking to demonstrate the conformity with this LVD Directive.*

<b><i>Manufacture's Name and Address:</i></b>	<i>ACCU-SCOPE Inc 73 Mall Drive Commack, New York 11725</i>
<b><i>Product Identification:</i></b>	<i>EXI-410 Series Microscopes EXI-410 Series Fluorescence Microscopes</i>
<b><i>Description:</i></b>	<i>Laboratory Clinical Research Microscope</i>
<b><i>Trade Name:</i></b>	<i>ACCU-SCOPE Inc</i>
<b><i>Model Number:</i></b>	<i>EXI-410 Series Microscopes</i>
<b><i>Identification of Test Standards:</i></b>	<i>EN 61010-1:2010</i>
<b><i>Technical Data:</i></b>	<i>100-240V, 50/60 Hz, 15W, IP20, Class I</i>
<b><i>Report No.:</i></b>	<i>EN180620010S</i>
<b><i>Testing Facilities:</i></b>	<i>EMTEK</i>

***The holder of this Certificate is authorized to use in connection with CE declaration of conformity with this LVD Directive.***

***Year of CE Markings: July 2018***

*Heston Singh  
VP, Business Development  
Legally binding signature of the issuer*